CERTIFICATE OF CONFORMITY

Certificate No.: CKEYS250509076

Applicant	:	**Jiaxing Dawnsail Biotech Co ., Ltd.**
Address	:	**Building 1, 501 Changshengnan Road, Jiaxing, Zhejiang, China**
Manufacturer	:	**Jiaxing Dawnsail Biotech Co ., Ltd.**
Address	:	**Building 1, 501 Changshengnan Road, Jiaxing, Zhejiang, China**
Trade mark	:	
Product	:	**mFIND nucleic acid analyzer**
Power Rating	:	**DC24V 6A**
		Adapter Type: ADP-96H24
		Input: 100-240V~, 50/60Hz, 1.5A MAX
		Output: DC24V 6A
Model No.	:	**mFIND-100**

The submitted sample of the above equipment has been tested and found to comply with the following European Directive:

Low Voltage Directive - 2014/35/EU

The standard(s) used for showing compliance with the essential requirements:

Applicable Standard(s) **EN 61010-1: 2010+A1:2019**

This certificate is part of the full test report(s) and should be read in conjunction with the test report: RKEYS250428087.
This certificate is based on an evaluation of one sample of above mentioned product. It does not imply assessment of the production of the product. Without the written approval of Guangdong KEYS Testing Technology Co., Ltd.this certificate is not permitted to be reproduced, except in full. It is not permitted to use the test lab's logo. The CE marking may only be used if all the relevant and effective European Directives are applicable.

CE

Guangdong KEYS Testing Technology Co., Ltd.
Jason Zhan
KEYS

Jason Zhan
Manager
Date: May. 09, 2025

Scan to view the original file

This report is only responsible for the test results of the samples submitted for inspection, and is not responsible for the source of the samples submitted for inspection.
This report shall not be altered, increased or deleted. Without written approval of KEYS, this test report shall not be copied except in full and published as advertisement.

Сертификат сертификации diplômes 認証証書 인증서 Сертификат сертификации diplômes 認証証書 인증서